

Prefilled Syringe Details

(Needle cover is removed from the needle)

(Syringe plunger is pushed in to demonstrate usage)

(Syringe is removed from the patient)

The safety device is designed to avoid accidental needle sticks and prohibit reuse of the syringe.

(Syringe tilts downwards)

After firmly pushing the plunger rod, an audible click signals that the needle guard is activated and ready for disposal.

(Safety system is activated)

(Picture pointing out Needle Cover)

(Picture pointing out Needle)

(Picture pointing out Plunger)

(Picture pointing out Safety Device)

[End of transcript]